

# METRALINE ISO<sup>CHECK</sup>

## Insulation Measuring Instrument

 3-349-690-03  
 4/5.21

- Insulation resistance measurement** with test voltages of 50 to 1000 V
- Voltage measurement** up to 600 V
- Measurement of surge protection devices** with test voltages of 50 to 1000 V
- Table of common varistors can be displayed
- Digital display, backlit color OLED display
- Indication of dangerous touch voltage
- LED for measurement point illumination
- Patented means of securing test probes
- Compact and rugged:**  
For service calls under harsh conditions and laboratory use



Ramp Function



Measurement Results



Table Display



### Applications

- Measurement of insulation resistance at voltage-free devices and systems, up to 1000 V depending upon variant
- Measurement of surge protection devices, up to 1000 V depending upon variant
- Checking of test objects for absence of voltage

### Applicable Regulations and Standards

IEC 61010-1/-031 DIN EN 61010-1/-031 VDE 0411-1/-031	Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements Part 31: Safety requirements for hand-held probe assemblies for electrical measurement and test
IEC 61557-1/-2 DIN EN 61557-1/-2 VDE 0413-1/-2	Electrical safety in low voltage distribution systems up to 1000 V AC and 1500 V DC – Equipment for testing, measuring or monitoring of protective measures Part 1: General requirements <b>Part 2: Insulation resistance measuring instruments</b>
IEC 61326-1 DIN EN 61326-1	Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1: General requirements
DIN EN 60529 VDE 0470-1	Degrees of protection provided by enclosures (IP code)

### Characteristic Values

#### Measurement of Insulation Resistance

Nominal Range per EN 61557-2: 0.100 MΩ – Rmax\*

Range	Reso-lution	Intrinsic Uncertainty	Measuring Uncertainty
0.100 to 9.999 MΩ	0.001 MΩ	(2% rdg. + 10 d)	(3 % rdg. + 20 d)
10.00 to 99.99 MΩ	0.01 MΩ	(2% rdg. + 10 d)	(3 % rdg. + 20 d)
100.0 to 999.9 MΩ	0.1 MΩ	(2% rdg. + 10 d)	(3 % rdg. + 20 d)
1.000 GΩ ... Rmax*	0.001 GΩ	(4 % rdg. + 15 d)	(5 % rdg. + 25 d)

\* The Rmax value depends on the selected test voltage:

Nominal voltage of 50 to 99 V      Rmax = 1.999 GΩ  
 Nominal voltage of 100 to 249 V      Rmax = 3.999 GΩ  
 Nominal voltage of 250 to 1000 V      Rmax = 9.999 GΩ

Nominal measuring voltage

50 to 1000 V  
 adjustable in steps of 1 V  
 -0%/+10% of nominal voltage

Measuring voltage

Nominal measuring current  
 Short-circuit current

50 to 1000 V  
 -0%/+10% of nominal voltage  
 ≥ 1 mA (where Umes > Unom)  
 < 3 mA

Nominal measuring current

Automatic discharging of the DUT  
 Number of measurements

50 to 1000 V  
 -0%/+10% of nominal voltage  
 ≥ 1 mA (where Umes > Unom)  
 < 3 mA

Short-circuit current

Yes

Automatic discharging of the DUT

approx. 250  
 (with new alkaline batteries)

# METRALINE ISO<sup>CHECK</sup>

## Insulation Measuring Instrument

### Measurement of Surge Protection Devices

Range	Resolution	Intrinsic Uncertainty	Measuring Uncertainty
40 to 1050 V	1 V	(2% rdg. + 2 d)	(3 % rdg. + 3 d)

Measuring Method Rising DC voltage when measuring the so-called milliamperes point

### Measurement of Direct and Alternating Voltage (Frequency Range: 45 to 60 Hz)

Range	Resolution	Intrinsic Uncertainty	Measuring Uncertainty
0 to 600 V	1 V	(2% rdg. + 2 d)	(3 % rdg. + 3 d)

### Key

- a) The TRMS value for alternating voltage is measured.
- b) rdg. means reading, i.e. measured value.
- d = digits (i.e. number of the decimal place with the least significance)

### Reference Conditions

Temperature	23 ± 2 °C
Relative humidity	40 to 60%
Device position	any

### Ambient Conditions

#### Operating Conditions

Operating temperature	0 to 40 °C
Relative humidity	max. 85 %, no condensation allowed
Device position	any

#### Storage Conditions

Temperature	-10 to 70 °C
Relative humidity	max. 90% at -10 to +40 °C max. 80% at +40 to +70 °C
Device position	any

### Power Supply

Batteries	4 ea. AAA (LR03), 1.5 V alkaline or 1.2 V NiMH (with at least 750 mAh)
Number of measurements	with batteries at 800 mAh: approx. 1,000 measurements (with 500 V test voltage on 500 kΩ)

### Electrical Safety

Measuring category	with safety cap applied to test probe: CAT III 300 V without safety cap applied to test probe: CAT II 300 V
Pollution degree	2
Protection class	II

### Electromagnetic Compatibility (EMC)

Interference emission EN 61326-1:2006 class B  
Interference immunity EN 61326-1:2006

### Mechanical Design

Display	OLED, multicolored, graphic
Protection	Housing: IP 43
Dimensions	approx. 260 x 70 x 40 mm
Weight	approx. 0.36 kg with batteries

### Scope of Delivery

- 1 Test instrument with mobile test probe
- 1 Pouch
- 1 CD ROM with operating instructions in available languages
- 1 Condensed operating instructions

### Order Information

Description	Type	Article number
Insulation measuring instrument	METRALINE ISOCHECK	M507C
Broad-range charger for charging optionally available batteries, e.g. Z507B, inserted in the METRALINE ISO-RCD-Z CHECK		
Input*: 100 to 240 V AC ±10%; Output: 9 V DC, 180 mA	Charger METRALINE CHECK Series	Z507A
4 rechargeable batteries (AAA) for METRALINE ISO-RCD-Z/CHECK	Akku-Set METRALINE CHECK Series	Z507B

\* with plug adapter for the following countries: EU, UK, US, AU

# Diensten van EURO-INDEX

**EURO-INDEX is fabrikant, importeur en distributeur van diverse A-merken op het gebied van test- en meetinstrumenten. Daarnaast leveren wij een groot aantal diensten om het gebruik van deze instrumenten in uw bedrijfsvoering te optimaliseren. Dit omvat uiteraard onderhoud, reparatie en kalibratie van de instrumenten, maar ook kennisdeling via de EURO-INDEX Academy en verhuur van instrumenten.**

## Geautoriseerd Service Centrum

EURO-INDEX b.v. is van alle vertegenwoordigde merken een Geautoriseerd Service Centrum. Dit betekent dat uw instrumenten worden behandeld door technici die zijn opgeleid door de fabrikant en beschikken over de juiste gereedschappen en software. Er worden uitsluitend originele onderdelen toegepast en de garantie van uw instrument, evenals de certificering (ATEX, EN50379, etc.) blijven intact.

## Kalibratielaboratorium

Ons moderne service- en kalibratielaboratorium beschikt over een RvA accreditatie naar NEN-EN-ISO/IEC 17025. Deze accreditatie geldt voor grootheden, zoals gespecificeerd in de scope bij accreditatienummer K105.



**Kijk voor een overzicht van al onze diensten op [euro-index.nl/diensten](http://euro-index.nl/diensten)**

## Mobiele Service

Naast de vaste kalibratielaboratoria in Capelle aan den IJssel en Zaventem beschikken wij ook over laboratoria op wielen met de naam "Mobiele Service". Dit biedt vertrouwde service en kwaliteit, bij u voor de deur!

## KWS®

KWS® is een uniek servicesysteem voor uw meetinstrumenten met periodiek onderhoud en kalibratie tegen vaste, lage kosten. Uw kalibratiecertificaten zijn digitaal beschikbaar via Mijn KWS (gratis webportaal en app).

## Verhuur van meetinstrumenten

- Uitgebreid assortiment
- Nauwkeurigheid aantoonbaar door actueel kalibratiecertificaat
- Deskundig advies
- Complete levering inclusief accessoires

## EURO-INDEX Academy

- Trainingen (individueel en klassikaal)
- Cursussen en workshops
- Demonstratie- en instructievideo's
- Whitepapers



Servicebalie



Onderhoud, reparatie en kalibratie



Cursussen en workshops



Mobiele Service

Wijzigingen voorbehouden EURO-INDEX® NL 24005



NETHERLAND

Rivium 2e straat 12  
2909 LG Capelle a/d IJssel  
T: 010 - 2 888 000  
F: 010 - 2 888 010  
verkoop@euro-index.nl  
[www.euro-index.nl](http://www.euro-index.nl)



Geaccrediteerde scope  
zie [www.rva.nl](http://www.rva.nl)



BELGIË

Leuvensesteenweg 607  
1930 Zaventem  
T: +32 - (0)2 - 757 92 44  
F: +32 - (0)2 - 757 92 64  
[info@euro-index.be](mailto:info@euro-index.be)  
[www.euro-index.be](http://www.euro-index.be)